



**Product No.:** PP1012

**Description:** Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

**Lot No.:** 35963

**Date of Manufacturing:** 12/4/2023

**Expiration Date:** 11/1/2025

**10% Formalin Lot #: 3k010**

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Clear liquid - Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)
pH @ 25°C	6.90 - 7.10	7.07

**Country of Origin:** United States of America

**Latex Free:** This product contains no latex

**Jonathan D. Kent, Ph.D.**  
**Director of Manufacturing**