



Product No.: PP1012

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

Lot No.: 36672

Date of Manufacturing: 2/5/2024

Expiration Date: 1/1/2026

10% Formalin Lot #: 4A015

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Clear liquid - Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.31% (v/v)
pH @ 25°C	6.90 - 7.10	6.95

Country of Origin: United States of America

Latex Free: This product contains no latex

Signature of Jonathan D. Kent, Ph.D., Director of Manufacturing