



Product No.: PP1012

Description: Kit, 12 place Prostate Biopsy, Cancer Diagnostics label, 12x20mL CDI vials half filled with 10% formalin per kit, 25 kits per case

Lot No.: 36932

Date of Manufacturing: 2/27/2024

Expiration Date: 2/1/2026

10% Formalin Lot #: 4B012

Specification	Requirements	Results
Appearance	Clear, colorless, no observable particulate	Clear liquid - Conforms
Formaldehyde Concentration	9.50 -10.50% (v/v)	10.00% (v/v)
pH @ 25°C	6.90 - 7.10	7.00

Country of Origin: United States of America

Latex Free: This product contains no latex

Signature of Jonathan D. Kent, Ph.D., Director of Manufacturing